

Recombinant allergen Ole e 1 for *Olea europaea* (olive tree)

CATALOG NUMBER: RAL0012

LOT NUMBER: #

RECOMBINANT ALLERGEN: *Olea europaea* ole e 1 olive tree allergen (Huecas *et al.*, 1999).

DESCRIPTION: Ole e 1 is the most prevalent allergen of olive tree pollen with more than ten isomeric forms. This recombinant allergen corresponds to the isoform 1.0102, and it has been prepared as a recombinant protein fused to a his-tag in its C-terminus.

PRESENTATION: liquid protein solution

SOURCE: *Pichia pastoris*

MOLECULAR WEIGHT: determined by SDS-PAGE, the protein is located between molecular markers of 18,400 and 25,000 Da, while relative molecular mass calculated from amino acid sequence, without glycosylation, is approx. 18,887.36 Da.

BATCH COMPOSITION:

COMPONENTS	COMPOSITION
his-ole e 1	recombinant allergen with a his-tag
Storage buffer	20 mM phosphate buffer pH 8 and 0.15 M NaCl

QUALITY CONTROL:

1. PROTEIN CONCENTRATION DETERMINED ESPECTROPHOTOMETRICALLY

DO₂₈₀ = 0.94

A_{0.1%} (=1 g/l) = 0.784

CONCENTRATION*: 1.20 mg/ml

* The measurement of the protein concentration has been performed with the theoretical extinction coefficient of the recombinant protein obtained from Gill and vonHippel, 1989. It is recommended that the users carry out their absorbance determinations to avoid equipment variabilities regarding final concentration, mainly in reproducibility analysis.

2. PURITY CONTROL IN SDS-PAGE: 17%





Figure 1. SDS-PAGE analysis (17%) of 6 µl of the recombinant allergen ole e 1. Purity is > 95% as determined by gel electrophoresis.

3. PROTEIN FINGERPRINT BY MASS SPECTROMETRY

Top Score: 117 for ole e 1

Protein score is $-10 \cdot \log(P)$, where P is the probability that the observed match is a random event. Protein scores greater than 13 are significant ($p < 0.05$).



P19963_2 Mass: 19217 Score: 117 Expect: 8e-012 Matches: 13
LOCUS ALL1_OLEEU

The MS was performed with a by MALDI TOF/TOF model UltrafleXtreme (Bruker).

4. ANALYSIS BY AN ELISA ASSAY

This biomarker has been evaluated in an external study carried out at a Spanish hospital by a group of allergists with positive and negative serum samples from patients. The evaluation of the recombinant allergens has been performed by means of an in-house ELISA assay. In this immunoassay, it has been determined the presence of specific IgE in sera that had previously been validated by skin prick testing (SPT) and the UniCAP® test. The sera panel for this study was composed of 25 positive and 10 negative specimen sera.

The recombinant allergen ole e 1 detected 25 positive sera out of 25 (100% incidence) with higher prevalence of sera with titers of 3.5-17.50 IU/ml (international units per milliliter; 1 IU is equivalent to 2.42 ng of IgE).

5. ABSENCE OF PRECIPITATION AFTER A FREEZING AND THAWING CYCLE: ok

LOT SPECIFICATIONS:

1. CONCENTRATION: 1.20 mg/ml

2. TOTAL QUANTITY PER ALIQUOT: 1 mg

3. TOTAL VOLUME PER ALIQUOT: 0.875 ml

4. STORAGE: Protein is shipped with dry ice. Upon arrival, it should be aliquoted to avoid repeated freezing and thawing cycles and stored at -20°C to -80°C. In order to defrost the protein, maintain the aliquot at 25°C without shaking to avoid aggregation.

5. TESTED APPLICATIONS: ELISA.

6. POSSIBLE APPLICATIONS: WB, DB, Indirect ELISA, positive control in direct ELISA, CLIA, lateral-flow. Where this product has not been tested for use in a particular technique, this does not necessarily exclude its use in such procedures. Suggested working dilutions are given as a guide only. It is recommended that the user titrates. **Recombinant allergens expressed in *E. coli* do not contain glycosylation, and *P. pastoris* adds a different glycosylation than the one found in the environment and food allergens. As a result, allergens produced in these expression systems do not need anti-CCD blocker to eliminate the anti-CCD IgE antibodies in normal human sera. In the specific case when you measure IgG with an allergen produced in *P. pastoris*, we recommend using our anti-CCD blocker (Ref. Rekom SOR0001) to avoid cross-reaction with antibodies in sera which reacts with fungal high mannose glycans.**

7. OBSERVATIONS: proteins should be maintained frozen at high concentrations. In order to defrost the protein, maintain the aliquot at 25°C without shaking to avoid aggregation. Prior making test dilutions and after defrosting the protein, is recommended to remove possible protein aggregates by centrifuging the stock solution, avoiding alterations in the immobilization of the biomolecule to the solid surface.

RELATED PRODUCTS:

Ole e 2, Ole e 5.

BIBLIOGRAPHY:

Huecas, S., Villalba, M., González, E., Martínez-Ruiz, A., and R., Rodríguez. Production and detailed characterization of biologically active olive pollen allergen Ole e 1 secreted by the yeast *Pichia pastoris*. 1999, Eur. J. Biochem. 261, 539-545.

Gill SC, von Hippel PH. Calculation of protein extinction coefficients from amino acid sequence data. *Anal Biochem*. 1989 Nov 1;182(2):319-26.

Important Notes: During shipment, small volumes of product will occasionally become entrapped in the seal of the product vial. For products with volumes of 200 µl or less, we recommend gently tapping the vial on a hard surface or briefly centrifuging the vial in a tabletop centrifuge to dislodge any liquid in the containers cap.

Although recombinant antigens are expressed in non-pathogenic *P. pastoris* and bacterial integrity is destroyed during purification, the antigen preparation should be handled as potentially infectious.

FOR RESEARCH AND COMMERCIAL USE IN VITRO: not for human in vivo or therapeutic use.